Annex D1

Calibration Certificates for Dust Monitoring Equipment



東業德勤測試顧問有限公司 ETS-TESTCONSULT LTD.

8/F Block B, Veristrong Industrial Centre 34-36 Au Pui Wan Street, Fo Tan, Hong Kong

T: +852 2695 8318 F: +852 2695 3944 E: etl@ets-testconsult.com W: www.ets-testconsult.com

Calibration Report of High Volume Air Sampler

Manufacturer

Graseby 105

Date of Calibration

12 April 2019

Serial No.

9795 (ET/EA/003/18)

Calibration Due Date

11 June 2019

Method


Five-point calibration by using standard calibration kit Tisch TE-5025A refer to the

Operations Manual

Results

Flow recorder reading (cfm)		48	43	36	31	27
Qstd (Actual flow rate, m³/min)		1.69	1.52	1.29	1.05	0.86
Pressure: 762.06 mm Hg		Temp. :	296	K .		

Sampler 9795 Calibration Curve Site: Tseung Kwan O 137 (TKO-A1)



Acceptance Criteria: Correlation coefficient (r) of the calibration curve greater than 0.990 after a 5-point calibration

The high volume sampler complies* / does not comply* with the specified requirements and is deemed acceptable*/ unacceptable* for use.

Calibrated by :

LIAO, Yun Chao (Technician)

Checked by

LAU, Chi Leung

(Environmental Team Leader)



東業德勤測試顧問有限公司 ETS-TESTCONSULT LTD.

8/F Block B, Veristrong Industrial Centre, 34-36 Au Pui Wan Street, Fo Tan, Hong Kong

T: +852 2695 8318 F: +852 2695 3944 E: etl@ets-testconsult.com W: www.ets-testconsult.com

Calibration Report of High Volume Air Sampler

Manufacturer

Andersen G1051

Date of Calibration

12 April 2019

Serial No.

1176 (ET/EA/003/05)

Calibration Due Date

11 June 2019

Method


Based on Operations Manual for the 5-point calibration using standard calibration kit

manufactured by Tisch TE-5025 A

Results

Flow recorder reading (cfm)		55	47	42	33	25
Qstd (Actual flow rate, m³/min)		1.75	1.52	1.37	1.01	0.83
Pressure :	762.06 mm Hg		Temp. :	296	K	

Sampler 1176 Calibration Curve Site: Tseung Kwan O 137 (TKO-A2a)



Acceptance Criteria: Correlation coefficient (r) of the calibration curve greater than 0.990 after a 5-point calibration

The high volume sampler complies* / does not comply* with the specified requirements and is deemed acceptable* / unacceptable * for use.

Calibrated by:

LIAÓ, Yun Chao (Technician) Checked by :

LAU, Chi Leung

(Environmental Team Leader)

- END OF REPORT -