


Annex D1

Calibration Certificates for Dust Monitoring Equipment




東業德勤測試顧問有限公司 ETS-TESTCONSULT LTD.™

8/F Block B,
Veristrong Industrial Centre,
34-36 Au Pui Wan Street,
Fo Tan, Hong Kong
T: +852 2695 8318
F: +852 2695 3944
E: eit@ets-testconsult.com
W: www.ets-testconsult.com

Calibration Report of High Volume Air Sampler

Manufacturer	:	Graseby 105	Date of Calibration	:	23 October 2019		
Serial No.	:	9795 (ET / EA / 003 / 18)	Calibration Due Date	:	22 December 2019		
Method	:	Five-point calibration by using standard calibration kit Tisch TE-5025A refer to the Operations Manual					
Results	:	Flow recorder reading (cfm)	49	42	38	31	27
		Qstd (Actual flow rate, m³/min)	1.69	1.51	1.27	1.02	0.83
		Pressure : 760.56 mm Hg		Temp. :	299 K		

Sampler 9795 Calibration Curve Site: Tseung Kwan O 137 (TKO-A1)




Acceptance Criteria : Correlation coefficient (r) of the calibration curve greater than 0.990 after a 5-point calibration

The high volume sampler complies* / does not comply* with the specified requirements and is deemed acceptable*/ unacceptable* for use.

Calibrated by : MAK, Kei Wai
MAK, Kei Wai
(Assistant Supervisor)

Checked by : LAU, Chi Leung
LAU, Chi Leung
(Environmental Team Leader)




東業德勤測試顧問有限公司 ETS-TESTCONSULT LTD.TM

8/F Block B,
Veristrong Industrial Centre,
34-36 Au Pui Wan Street,
Fo Tan, Hong Kong
T: +852 2695 8318
F: +852 2695 3944
E: etl@ets-testconsult.com
W: www.ets-testconsult.com

Calibration Report of High Volume Air Sampler

Manufacturer	:	Andersen G1051	Date of Calibration	:	23 October 2019
Serial No.	:	1176 (ET / EA / 003 / 05)	Calibration Due Date	:	22 December 2019
Method	:	Based on Operations Manual for the 5-point calibration using standard calibration kit manufactured by Tisch TE-5025 A			
Results	:	Flow recorder reading (cfm)	55	48	45
		Qstd (Actual flow rate, m ³ /min)	1.74	1.52	1.37
		Pressure :	760.56 mm Hg	Temp. :	299 K

Sampler 1176 Calibration Curve
Site: Tseung Kwan O 137 (TKO-A2a)



Acceptance Criteria : Correlation coefficient (r) of the calibration curve greater than 0.990 after a 5-point calibration

The high volume sampler complies* / does not comply* with the specified requirements and is deemed acceptable* / unacceptable * for use.

Calibrated by : Mark Kei Wai

MAK, Kei Wai
(Assistant Supervisor)

Checked by : Chi Leung

LAU, Chi Leung
(Environmental Team Leader)