


Annex D1

# Calibration Certificates for Dust Monitoring Equipment









**Calibration Report**  
of  
**High Volume Air Sampler**


Manufacturer : Andersen G1051 Date of Calibration : 07 August 2020

Serial No. : 1176 ( ET / EA / 003 / 05 ) Calibration Due Date : 06 October 2020

Method : Based on Operations Manual for the 5-point calibration using standard calibration kit manufactured by Tisch TE-5025 A

Results	Flow recorder reading (cfm)	54	46	41	34	27
	Qstd (Actual flow rate, m <sup>3</sup> /min)	1.83	1.64	1.50	1.17	0.92
	Pressure :	753.81 mm Hg			Temp. :	306 K

**Sampler 1176 Calibration Curve**  
Site: Tseung Kwan O 137 (TKO-A2a)



Acceptance Criteria : Correlation coefficient (r) of the calibration curve greater than 0.990 after a 5-point calibration

The high volume sampler complies\* / does not comply\* with the specified requirements and is deemed acceptable\* / unacceptable\* for use.

Calibrated by :   
CHAN, Wai Man  
(Technician)

Checked by :   
LAU, Chi Leung  
(Environmental Team Leader)