

Annex D1

Calibration Certificates for Dust Monitoring Equipment

TSP SAMPLER CALIBRATION CALCULATION SPREADSHEET

Location ID : AM1	Date of Calibration: 24-Nov-21
Name and Model: TISCH HVS Model TE-5170	Next Calibration Date: 24-Jan-22
	Technician: Fai So

CONDITIONS

Sea Level Pressure (hPa)	1020.3	Corrected Pressure (mm Hg)	765.225
Temperature (°C)	19.0	Temperature (K)	292

CALIBRATION ORIFICE

Make-> TISCH	Qstd Slope -> 2.10574
Model-> 5025A	Qstd Intercept -> -0.00985
Serial # -> 1941	

CALIBRATION

Plate No.	H2O (L) (in)	H2O (R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC corrected	LINEAR REGRESSION
18	6.20	6.20	12.4	1.700	59	60.42	Slope = 37.2995 Intercept = -2.4242 Corr. coeff. = 0.9993
13	4.70	4.70	9.4	1.481	52	53.25	
10	3.70	3.70	7.4	1.314	46	47.11	
7	2.40	2.40	4.8	1.059	36	36.87	
5	1.50	1.50	3.0	0.838	28	28.67	

Calculations :

$$Qstd = 1/m[\text{Sqrt}(H2O(Pa/Pstd)(Tstd/Ta))]-b$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart responses

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pstd = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$


m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure



TSP SAMPLER CALIBRATION CALCULATION SPREADSHEET

Location ID : AM2	Date of Calibration: 24-Nov-21
Name and Model: TISCH HVS Model TE-5170	Next Calibration Date: 24-Jan-22
	Technician: Fai So

CONDITIONS

Sea Level Pressure (hPa) 1020.3	Corrected Pressure (mm Hg) 765.225
Temperature (°C) 19.0	Temperature (K) 292

CALIBRATION ORIFICE

Make-> TISCH	Qstd Slope -> 2.10574
Model-> 5025A	Qstd Intercept -> -0.00985
Serial # -> 1941	

CALIBRATION

Plate No.	H2O (L) (in)	H2O (R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC corrected	LINEAR REGRESSION		
							Slope =	Intercept =	Corr. coeff. =
18	7.20	7.20	14.4	1.831	51	52.23	Slope = 30.7872 Intercept = -3.3292 Corr. coeff. = 0.9980		
13	5.50	5.50	11.0	1.601	46	47.11			
10	4.40	4.40	8.8	1.433	40	40.96			
7	2.70	2.70	5.4	1.123	30	30.72			
5	1.50	1.50	3.0	0.838	22	22.53			

Calculations :

$$Qstd = 1/m[\text{Sqrt}(H2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart responses

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pstd = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$


m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure



TSP SAMPLER CALIBRATION CALCULATION SPREADSHEET

Location ID : AM3	Date of Calibration: 24-Nov-21
Name and Model: TISCH HVS Model TE-5170	Next Calibration Date: 24-Jan-22
	Technician: Fai So

CONDITIONS

Sea Level Pressure (hPa) 1020.3	Corrected Pressure (mm Hg) 765.225
Temperature (°C) 19.0	Temperature (K) 292

CALIBRATION ORIFICE

Make-> TISCH	Qstd Slope -> 2.10574
Model-> 5025A	Qstd Intercept -> -0.00985
Serial # -> 1941	

CALIBRATION

Plate No.	H2O (L) (in)	H2O (R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC corrected	LINEAR REGRESSION		
							Slope =	Intercept =	Corr. coeff. =
18	6.80	6.80	13.6	1.780	54	55.30	Slope = 35.7467 Intercept = -6.9119 Corr. coeff. = 0.9944		
13	5.50	5.50	11.0	1.601	50	51.20			
10	4.10	4.10	8.2	1.383	42	43.01			
7	2.70	2.70	5.4	1.123	34	34.82			
5	1.60	1.60	3.2	0.866	22	22.53			

Calculations :

$$Qstd = 1/m[\text{Sqrt}(H2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart responses

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K

Pstd = actual pressure during calibration (mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$


m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure



TSP SAMPLER CALIBRATION CALCULATION SPREADSHEET

Location ID : AM4	Date of Calibration: 24-Nov-21
Name and Model: TISCH HVS Model TE-5170	Next Calibration Date: 24-Jan-22
	Technician: Fai So

CONDITIONS

Sea Level Pressure (hPa) 1020.3	Corrected Pressure (mm Hg) 765.225
Temperature (°C) 19.0	Temperature (K) 292

CALIBRATION ORIFICE

Make-> TISCH	Qstd Slope -> 2.10574
Model-> 5025A	Qstd Intercept -> -0.00985
Serial # -> 1941	

CALIBRATION

Plate No.	H2O (L) (in)	H2O (R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC corrected	LINEAR REGRESSION		
							Slope =	Intercept =	Corr. coeff. =
18	6.30	6.30	12.6	1.713	49	50.18	Slope = 30.9360 Intercept = -2.2579 Corr. coeff. = 0.9984		
13	5.20	5.20	10.4	1.557	45	46.08			
10	3.80	3.80	7.6	1.332	39	39.94			
7	2.50	2.50	5.0	1.081	30	30.72			
5	1.50	1.50	3.0	0.838	23	23.55			

Calculations :

$$Qstd = 1/m[\text{Sqrt}(H2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart responses

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pstd = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$

m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

