Annex F1

Calibration Certificates for Surface Water Quality Monitoring Equipment



ALS Technichem (HK) Pty Ltd 11/F., Chung Shun Knitting Centre, 1 - 3 Wing Yip Street, Kwai Chung, N.T., Hong Kong **T:** +852 2610 1044 **F:** +852 2610 2021 www.alsglobal.com

# **REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION**

| CONTACT:<br>CLIENT: | MR IVAN LEUNG<br>ALS TECHNICHEM (HK) PTY LTD | WORK ORDER:    | HK2300781   |
|---------------------|--|----------------|-------------|
| ADDRESS:            | 11/F., CHUNG SHUN KNITTING CENTRE,           | SUB-BATCH:     | 0           |
|                     | 1-3 WING YIP STREET, KWAI CHUNG, N.T.        | LABORATORY:    | HONG KONG   |
|                     |  | DATE RECEIVED: | 05-Jan-2023 |
|                     |  | DATE OF ISSUE: | 18-Jan-2023 |

### SPECIFIC COMMENTS

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client. The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the

laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

| Equipment Type:            | Multifunctional Meter                            |
|----------------------------|--|
| Service Nature:            | Performance Check                                |
| Scope:                     | Conductivity, pH Value, Salinity and Temperature |
| Brand Name/ Model No.:     | [YSI]/ [Professional Plus]                       |
| Serial No./ Equipment No.: | [15G100349/JC024046]/ [HK1274]                   |
| Date of Calibration:       | 11-January-2023                                  |

### **GENERAL COMMENTS**

This report superseded any previous report(s) with same work order number.

Ms. Lin Wai Yu, Iris Assistant Manager - Inorganics

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

# **REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION**



| WORK ORDER:                             | HK2300781                                       |                           |               |
|---|---|---------------------------|---------------|
| SUB-BATCH:<br>DATE OF ISSUE:<br>CLIENT: | 0<br>18-Jan-2023<br>ALS TECHNICHEM (HK) PTY LTD |                           |               |
| Equipment Type:                         | Multifunctional Meter                           |                           |               |
| Brand Name/<br>Model No.:               | [YSI]/ [Professional Plus]                      |                           |               |
| Serial No./<br>Equipment No.:           | [15G100349/JC024046]/ [HK127                    | 74]                       |               |
| Date of Calibration:                    | 11-January-2023                                 | Date of Next Calibration: | 11-April-2023 |

#### PARAMETERS:

#### Conductivity Method Ref: APHA (23rd edition), 2510B

| Expected Reading (µS/cm) | Displayed Reading (µS/cm) | Tolerance (%) |
|--------------------------|---------------------------|---------------|
| 146.9                    | 143.5                     | -2.3          |
| 6667                     | 6637                      | -0.4          |
| 12890                    | 12868                     | -0.2          |
| 58670                    | 57072                     | -2.7          |
|                          | Tolerance Limit (%)       | ±10.0         |

#### pH Value

#### Method Ref: APHA (23rd edition), 4500H: B

| Expected Reading (pH unit) | Displayed Reading (pH unit) | Tolerance (pH unit) |  |
|----------------------------|-----------------------------|---------------------|--|
| 4.0                        | 4.05                        | +0.05               |  |
| 7.0                        | 7.15                        | +0.15               |  |
| 10.0                       | 10.04                       | +0.04               |  |
|                            | Tolerance Limit (pH unit)   | ±0.20               |  |

#### Salinity

#### Method Ref: APHA (23rd edition), 2520B

| Expected Reading (ppt) | Displayed Reading (ppt) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0                      | 0.00                    |               |
| 10                     | 9.53                    | -4.7          |
| 20                     | 19.77                   | -1.2          |
| 30                     | 29.15                   | -2.8          |
|                        | Tolerance Limit (%)     | ±10.0         |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris Assistant Manager - Inorganics

## **REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION**

| LS) |
|-----|

| WORK ORDER:   | HK2300781   |                           |               |
|---|---|---------------------------|---------------|
| SUB-BATCH:<br>DATE OF ISSUE:<br>CLIENT:                     | 0<br>18-Jan-2023<br>ALS TECHNICHEM (HK) PTY LTD                                     | ,                         |               |
| Equipment Type:<br>Brand Name/<br>Model No.:<br>Serial No./ | Multifunctional Meter<br>[YSI]/ [Professional Plus]<br>[15G100349/JC024046]/ [HK123 | 74]                       |               |
| Equipment No.:<br>Date of Calibration:                      | 11-January-2023   | Date of Next Calibration: | 11-April-2023 |

#### **PARAMETERS:**

#### Temperature

#### Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Expected Reading (°C) | Displayed Reading (°C) | Tolerance (°C) |
|-----------------------|------------------------|----------------|
| 6.5                   | 5.7                    | -0.8           |
| 21.5                  | 21.2                   | -0.3           |
| 42.5                  | 42.4                   | -0.1           |
|                       | Tolerance Limit (°C)   | ±2.0           |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris Assistant Manager - Inorganics