



ANNEX D1

CALIBRATION CERTIFICATES FOR DUST
MONITORING EQUIPMENT

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM1 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 20-Feb-24 Next Calibration Date: 19-Apr-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1017</td></tr> <tr><td>24.0</td></tr> </table>	1017	24.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>762.8</td></tr> <tr><td>297</td></tr> </table>	762.8	297	
1017								
24.0								
762.8								
297								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.2	6.2	12.4	1.719	56	56.21	Slope= 30.370	
13	4.9	4.9	9.8	1.530	51	51.19	Intercept= 4.548	
10	3.6	3.6	7.2	1.313	45	45.17	Corr. Coeff.= 0.9989	
7	2.4	2.4	4.8	1.075	37	37.14		
5	1.5	1.5	3.0	0.853	30	30.11		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

Q_{std} = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Q_{std} slope
 b = calibrator Q_{std} intercept
 T_a = actual temperature during calibration (deg K)
 P_a = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$1/m((I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$

m = sampler slope
 b = sampler intercept
 I = chart response
 T_{avg} = daily average temperature
 P_{avg} = daily average pressure

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM2 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 20-Feb-24 Next Calibration Date: 19-Apr-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1017</td></tr> <tr><td>24.0</td></tr> </table>	1017	24.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>762.8</td></tr> <tr><td>297</td></tr> </table>	762.8	297	
1017								
24.0								
762.8								
297								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.5	6.5	13	1.759	52	52.20	Slope= 26.514	
13	5.0	5.0	10	1.545	47	47.18	Intercept= 6.297	
10	3.6	3.6	7.2	1.313	42	42.16	Corr. Coeff.= 0.9933	
7	2.2	2.3	4.5	1.041	35	35.13		
5	1.4	1.5	2.9	0.839	27	27.10		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(Pa/Pstd)(Tstd/Ta)} - b]$

$IC = I[\sqrt{Pa/Pstd}(Tstd/Ta)]$

Qstd = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Qstd slope
 b = calibrator Qstd intercept
 Ta = actual temperature during calibration (deg K)
 Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:
 $1/m((I)[\sqrt{298/Tav}(Pav/760)] - b)$


m = sampler slope
 b = sampler intercept
 I = chart response
 Tav = daily average temperature
 Pav = daily average pressure

IC

Flow Rate

10
15
20
25
30
35
40
45
50
55
60

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM3 Name and Model : TISCH HVS Model TE-5170		Date of Calibration: 20-Feb-24 Next Calibration Date: 19-Apr-24 Operator: P.F.Yeung													
CONDITIONS															
Sea Level Pressure (hpa) Temperature (°C)	1017 24.0	Corrected Pressure (mm Hg) Temperature (K)	762.8 297												
CALIBRATION ORIFICE															
Make: Model: Serial#:	TISCH TE-5025A 2454	Qstd Slope Qstd Intercept	2.07544 -0.03205												
CALIBRATION															
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m³/min)	I (chart)	IC (corrected)	LINEAR REGRESSION								
18	5.6	5.7	11.3	1.641	57	57.22	Slope= 27.283								
13	4.3	4.3	8.6	1.434	52	52.20	Intercept= 12.645								
10	3.3	3.2	6.5	1.249	46	46.17	Corr. Coeff.= 0.9985								
7	2.0	2.0	4.0	0.983	40	40.15									
5	1.2	1.2	2.4	0.765	33	33.12									
Calulations: $Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$ $IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$ Qstd = standard flow rate IC = corrected chart response I = actual chart response m = calibrator Qstd slope b = calibrator Qstd intercept Ta = actual temperature during calibration (deg K) Pa = actual pressure during calibration (mm Hg)			 <table border="1"> <thead> <tr> <th>Qstd (m³/min)</th> <th>IC</th> </tr> </thead> <tbody> <tr><td>0.765</td><td>33.12</td></tr> <tr><td>1.000</td><td>40.15</td></tr> <tr><td>1.249</td><td>46.17</td></tr> <tr><td>1.434</td><td>52.20</td></tr> <tr><td>1.641</td><td>57.22</td></tr> </tbody> </table>	Qstd (m³/min)	IC	0.765	33.12	1.000	40.15	1.249	46.17	1.434	52.20	1.641	57.22
Qstd (m³/min)	IC														
0.765	33.12														
1.000	40.15														
1.249	46.17														
1.434	52.20														
1.641	57.22														
For subsequent calculation of sampler flow: $1/m((I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$ m = sampler slope b = sampler intercept I = chart response Tavg = daily average temperature Pav = daily average pressure															

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM4 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 20-Feb-24 Next Calibration Date: 19-Apr-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1017</td></tr> <tr><td>24.0</td></tr> </table>	1017	24.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>762.8</td></tr> <tr><td>297</td></tr> </table>	762.8	297	
1017								
24.0								
762.8								
297								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.0	6.0	12.0	1.691	57	57.22	Slope= 32.433	
13	4.7	4.8	9.5	1.506	51	51.19	Intercept= 1.905	
10	3.7	3.7	7.4	1.331	44	44.17	Corr. Coeff.= 0.9966	
7	2.4	2.5	4.9	1.086	36	36.14		
5	1.5	1.4	2.9	0.839	30	30.11		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

Qstd = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Qstd slope
 b = calibrator Qstd intercept
 Ta = actual temperature during calibration (deg K)
 Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:
 $1/m((I)[\sqrt{298/T_{avg}}(\bar{P}_a/760)] - b)$

m = sampler slope
 b = sampler intercept
 I = chart response
 Tavg = daily average temperature
 Pav = daily average pressure

Flow Rate

Qstd (m³/min)	IC
0.839	30.11
1.086	36.14
1.331	44.17
1.506	51.19
1.691	57.22