




ANNEX D1

CALIBRATION CERTIFICATES FOR DUST
MONITORING EQUIPMENT

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM1 Name and Model : TISCH HVS Model TE-5170		Date of Calibration: 22-Apr-24 Next Calibration Date: 22-Jun-24 Operator: P.F.Yeung					
CONDITIONS							
Temperature (°C)	1014 25.0	Corrected Pressure (mm Hg) Temperature (K)	760.1 298				
CALIBRATION ORIFICE							
Make:	TISCH	Qstd Slope	2.07544				
Model:	TE-5025A	Qstd Intercept	-0.03205				
Serial#:	2454						
CALIBRATION							
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m³/min)	I (chart)	IC (corrected)	LINEAR REGRESSION
18	5.4	5.4	10.8	1.600	52	52.03	Slope= 33.386
13	4.4	4.4	8.8	1.446	47	47.03	Intercept= -0.9831
10	2.9	3.0	5.9	1.187	40	40.02	Corr. Coeff.= 0.9974
7	2.2	2.2	4.4	1.027	33	33.02	
5	1.3	1.3	2.6	0.793	25	25.02	
Calulations:							
$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$							
$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$							
Q_{std} = standard flow rate							
IC = corrected chart response							
I = actual chart response							
m = calibrator Q_{std} slope							
b = calibrator Q_{std} intercept							
T_a = actual temperature during calibration (deg K)							
P_a = actual pressure during calibration (mm Hg)							
For subsequent calculation of sampler flow:							
$1/m(I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b$							
m = sampler slope							
b = sampler intercept							
I = chart response							
T_{avg} = daily average temperature							
P_{avg} = daily average pressure							



TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM2 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 22-Apr-24 Next Calibration Date: 22-Jun-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1014</td></tr> <tr><td>25.0</td></tr> </table>	1014	25.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>760.1</td></tr> <tr><td>298</td></tr> </table>	760.1	298	
1014								
25.0								
760.1								
298								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	5.0	5.0	10.0	1.540	52	52.03	Slope= 30.764	
13	4.3	4.3	8.6	1.429	47	47.03	Intercept= 3.559	
10	3.5	3.5	7.0	1.291	42	42.03	Corr. Coeff.= 0.9953	
7	2.1	2.1	4.2	1.004	35	35.02		
5	1.3	1.3	2.6	0.793	28	28.02		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(Pa/Pstd)(Tstd/Ta)} - b]$

$IC = I[\sqrt{Pa/Pstd}(Tstd/Ta)]$

Q_{std} = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Q_{std} slope
 b = calibrator Q_{std} intercept
 Ta = actual temperature during calibration (deg K)
 Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$1/m(I)[\sqrt{(298/Tav)(Pav/760)}] - b$

m = sampler slope
 b = sampler intercept
 I = chart response
 Tav = daily average temperature
 Pav = daily average pressure

Flow Rate

Qstd (m ³ /min)	IC
0.8	28
1.0	35
1.3	42
1.4	47
1.55	52

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM3 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 22-Apr-24 Next Calibration Date: 23-Jun-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1014</td></tr> <tr><td>25.0</td></tr> </table>	1014	25.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>760.1</td></tr> <tr><td>298</td></tr> </table>	760.1	298	
1014								
25.0								
760.1								
298								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	5.5	5.5	11.0	1.614	57	57.04	Slope= 26.552	
13	4.2	4.2	8.4	1.413	52	52.03	Intercept= 14.177	
10	3.4	3.4	6.8	1.273	47	47.03	Corr. Coeff.= 0.9952	
7	2.1	2.1	4.2	1.004	42	42.03		
5	1.3	1.4	2.7	0.808	35	35.02		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$


$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

Q_{std} = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Q_{std} slope
 b = calibrator Q_{std} intercept
 T_a = actual temperature during calibration (deg K)
 P_a = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$1/m(I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b$

m = sampler slope
 b = sampler intercept
 I = chart response
 T_{avg} = daily average temperature
 P_{avg} = daily average pressure



TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM4 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 22-Apr-24 Next Calibration Date: 23-Jun-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1014</td></tr> <tr><td>25.0</td></tr> </table>	1014	25.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>760.1</td></tr> <tr><td>298</td></tr> </table>	760.1	298	
1014								
25.0								
760.1								
298								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.5	6.4	12.9	1.747	59	59.04	Slope= 32.028	
13	5.1	5.2	10.3	1.563	52	52.03	Intercept= 2.6869	
10	3.8	3.9	7.7	1.353	46	46.03	Corr. Coeff.= 0.9989	
7	2.5	2.4	4.9	1.083	38	38.02		
5	1.6	1.5	3.1	0.864	30	30.02		

Calulations:

Qstd = $1/m[\sqrt{H2O(Pa/Pstd)(Tstd/Ta)} - b]$
 IC = $I[\sqrt{Pa/Pstd}(Tstd/Ta)]$

Qstd = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Qstd slope
 b = calibrator Qstd intercept
 Ta = actual temperature during calibration (deg K)
 Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:
 $1/m(I)[\sqrt{(298/Tav)(Pav/760)}] - b$

m = sampler slope
 b = sampler intercept
 I = chart response
 Tav = daily average temperature
 Pav = daily average pressure

IC **Flow Rate**

Qstd (m ³ /min)	IC
0.864	30.02
1.083	38.02
1.353	46.03
1.563	52.03
1.747	59.04

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM1 Name and Model : TISCH HVS Model TE-5170		Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung					
CONDITIONS							
Temperature (°C)	1005 32.0	Corrected Pressure (mm Hg) Temperature (K)	753.8 305				
CALIBRATION ORIFICE							
Make:	TISCH	Qstd Slope	2.07544				
Model:	TE-5025A	Qstd Intercept	-0.03205				
Serial#:	2454						
CALIBRATION							
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION
18	5.8	5.8	11.6	1.631	53	52.19	Slope= 33.376
13	4.5	4.5	9.0	1.439	48	47.26	Intercept= -1.264
10	3.4	3.3	6.7	1.244	42	41.36	Corr. Coeff.= 0.9971
7	2.2	2.2	4.4	1.011	33	32.49	
5	1.4	1.3	2.7	0.795	25	24.62	
Calulations:							
$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$							
$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$							
Q_{std} = standard flow rate							
IC = corrected chart response							
I = actual chart response							
m = calibrator Q_{std} slope							
b = calibrator Q_{std} intercept							
T_a = actual temperature during calibration (deg K)							
P_a = actual pressure during calibration (mm Hg)							
For subsequent calculation of sampler flow:							
$1/m(I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$							
m = sampler slope							
b = sampler intercept							
I = chart response							
T_{avg} = daily average temperature							
P_{avg} = daily average pressure							

Flow Rate

Qstd (m3/min)	IC
0.8	25
1.0	33
1.2	42
1.6	53

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM2 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>32.0</td></tr> </table>	1005	32.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>305</td></tr> </table>	753.8	305	
1005								
32.0								
753.8								
305								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.1	6.1	12.2	1.673	53	52.19	Slope= 28.030	
13	4.5	4.5	9.0	1.439	48	47.26	Intercept= 6.135	
10	3.6	3.6	7.2	1.289	43	42.34	Corr. Coeff.= 0.9951	
7	2.2	2.2	4.4	1.011	36	35.45		
5	1.5	1.5	3.0	0.837	29	28.56		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

Qstd = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Qstd slope
 b = calibrator Qstd intercept
 Ta = actual temperature during calibration (deg K)
 Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:
 $1/m((I)[\sqrt{298/T_{avg}}(P_{avg}/760)] - b)$

m = sampler slope
 b = sampler intercept
 I = chart response
 Tavg = daily average temperature
 Pav = daily average pressure

Flow Rate

Qstd (m³/min)	IC
0.83	28.56
1.01	35.45
1.30	42.34
1.45	47.26

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM3 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>32.0</td></tr> </table>	1005	32.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>305</td></tr> </table>	753.8	305	
1005								
32.0								
753.8								
305								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	5.5	5.6	11.1	1.596	58	57.11	Slope= 27.386	
13	4.3	4.4	8.7	1.415	52	51.20	Intercept= 13.224	
10	3.0	3.0	6.0	1.178	47	46.28	Corr. Coeff.= 0.9981	
7	1.9	2.0	3.9	0.952	40	39.39		
5	1.2	1.2	2.4	0.750	34	33.48		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$


Q_{std} = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Q_{std} slope
 b = calibrator Q_{std} intercept
 T_a = actual temperature during calibration (deg K)
 P_a = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$1/m((I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$

m = sampler slope
 b = sampler intercept
 I = chart response
 T_{avg} = daily average temperature
 P_{avg} = daily average pressure

IC
Flow Rate



Qstd (m ³ /min)	Flow Rate
0.78	33
1.22	46
1.42	51
1.62	57

TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM4 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung							
CONDITIONS								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>32.0</td></tr> </table>	1005	32.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>305</td></tr> </table>	753.8	305	
1005								
32.0								
753.8								
305								
CALIBRATION ORIFICE								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
CALIBRATION								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m ³ /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.3	6.3	12.6	1.700	58	57.11	Slope= 33.491	
13	5.0	5.0	10	1.516	52	51.20	Intercept= 0.132	
10	3.7	3.8	7.5	1.315	44	43.33	Corr. Coeff.= 0.9989	
7	2.5	2.4	4.9	1.066	37	36.43		
5	1.5	1.4	2.9	0.823	28	27.57		

Calulations:

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

Q_{std} = standard flow rate
 IC = corrected chart response
 I = actual chart response
 m = calibrator Q_{std} slope
 b = calibrator Q_{std} intercept
 T_a = actual temperature during calibration (deg K)
 P_a = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$1/m((I)[\sqrt{298/T_{avg}}(P_{avg}/760)] - b)$

m = sampler slope
 b = sampler intercept
 I = chart response
 T_{avg} = daily average temperature
 P_{avg} = daily average pressure