



ANNEX D1

CALIBRATION CERTIFICATES FOR DUST  
MONITORING EQUIPMENT

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM1 Name and Model : TISCH HVS Model TE-5170		Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung					
<b>CONDITIONS</b>							
Temperature (°C)	1005 32.0	Corrected Pressure (mm Hg) Temperature (K)	753.8 305				
<b>CALIBRATION ORIFICE</b>							
Make:	TISCH	Qstd Slope	2.07544				
Model:	TE-5025A	Qstd Intercept	-0.03205				
Serial#:	2454						
<b>CALIBRATION</b>							
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION
18	5.8	5.8	11.6	1.631	53	52.19	Slope= 33.376
13	4.5	4.5	9.0	1.439	48	47.26	Intercept= -1.264
10	3.4	3.3	6.7	1.244	42	41.36	Corr. Coeff.= 0.9971
7	2.2	2.2	4.4	1.011	33	32.49	
5	1.4	1.3	2.7	0.795	25	24.62	
<b>Calulations:</b>							
$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$							
$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$							
$Q_{std}$ = standard flow rate							
$IC$ = corrected chart response							
$I$ = actual chart response							
$m$ = calibrator $Q_{std}$ slope							
$b$ = calibrator $Q_{std}$ intercept							
$T_a$ = actual temperature during calibration (deg K)							
$P_a$ = actual pressure during calibration (mm Hg)							
<b>For subsequent calculation of sampler flow:</b>							
$1/m(I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$							
$m$ = sampler slope							
$b$ = sampler intercept							
$I$ = chart response							
$T_{avg}$ = daily average temperature							
$P_{avg}$ = daily average pressure							

**Flow Rate**

Qstd (m3/min)	IC
0.8	25
1.0	33
1.2	42
1.6	53

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM2 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung							
<b>CONDITIONS</b>								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>32.0</td></tr> </table>	1005	32.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>305</td></tr> </table>	753.8	305	
1005								
32.0								
753.8								
305								
<b>CALIBRATION ORIFICE</b>								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
<b>CALIBRATION</b>								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.1	6.1	12.2	1.673	53	52.19	Slope= 28.030	
13	4.5	4.5	9.0	1.439	48	47.26	Intercept= 6.135	
10	3.6	3.6	7.2	1.289	43	42.34	Corr. Coeff.= 0.9951	
7	2.2	2.2	4.4	1.011	36	35.45		
5	1.5	1.5	3.0	0.837	29	28.56		

**Calulations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

Qstd = standard flow rate  
 IC = corrected chart response  
 I = actual chart response  
 m = calibrator Qstd slope  
 b = calibrator Qstd intercept  
 Ta = actual temperature during calibration (deg K)  
 Pa = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:  
 $1/m((I)[\sqrt{298/T_{avg}}(P_{avg}/760)] - b)$

m = sampler slope  
 b = sampler intercept  
 I = chart response  
 Tavg = daily average temperature  
 Pav = daily average pressure

**Flow Rate**

Qstd (m³/min)	IC
0.83	28.56
1.01	35.45
1.30	42.34
1.45	47.26

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM3 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung							
<b>CONDITIONS</b>								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>32.0</td></tr> </table>	1005	32.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>305</td></tr> </table>	753.8	305	
1005								
32.0								
753.8								
305								
<b>CALIBRATION ORIFICE</b>								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
<b>CALIBRATION</b>								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m <sup>3</sup> /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	5.5	5.6	11.1	1.596	58	57.11	Slope= 27.386	
13	4.3	4.4	8.7	1.415	52	51.20	Intercept= 13.224	
10	3.0	3.0	6.0	1.178	47	46.28	Corr. Coeff.= 0.9981	
7	1.9	2.0	3.9	0.952	40	39.39		
5	1.2	1.2	2.4	0.750	34	33.48		

**Calulations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$


$Q_{std}$  = standard flow rate  
 $IC$  = corrected chart response  
 $I$  = actual chart response  
 $m$  = calibrator  $Q_{std}$  slope  
 $b$  = calibrator  $Q_{std}$  intercept  
 $T_a$  = actual temperature during calibration (deg K)  
 $P_a$  = actual pressure during calibration (mm Hg)

For subsequent calculation of sampler flow:

$1/m((I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$

$m$  = sampler slope  
 $b$  = sampler intercept  
 $I$  = chart response  
 $T_{avg}$  = daily average temperature  
 $P_{avg}$  = daily average pressure

IC
Flow Rate



Qstd (m <sup>3</sup> /min)	Flow Rate
0.78	33
1.22	46
1.42	51
1.62	57

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM4 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Jun-24 Next Calibration Date: 19-Aug-24 Operator: P.F.Yeung							
<b>CONDITIONS</b>								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>32.0</td></tr> </table>	1005	32.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>305</td></tr> </table>	753.8	305	
1005								
32.0								
753.8								
305								
<b>CALIBRATION ORIFICE</b>								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
<b>CALIBRATION</b>								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m <sup>3</sup> /min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.3	6.3	12.6	1.700	58	57.11	Slope= 33.491	
13	5.0	5.0	10	1.516	52	51.20	Intercept= 0.132	
10	3.7	3.8	7.5	1.315	44	43.33	Corr. Coeff.= 0.9989	
7	2.5	2.4	4.9	1.066	37	36.43		
5	1.5	1.4	2.9	0.823	28	27.57		

**Calulations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

$Q_{std}$  = standard flow rate  
 $IC$  = corrected chart response  
 $I$  = actual chart response  
 $m$  = calibrator  $Q_{std}$  slope  
 $b$  = calibrator  $Q_{std}$  intercept  
 $T_a$  = actual temperature during calibration (deg K)  
 $P_a$  = actual pressure during calibration (mm Hg)

**For subsequent calculation of sampler flow:**

$1/m((I)[\sqrt{298/T_{avg}}(P_{avg}/760)] - b)$

$m$  = sampler slope  
 $b$  = sampler intercept  
 $I$  = chart response  
 $T_{avg}$  = daily average temperature  
 $P_{avg}$  = daily average pressure

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM1 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Aug-24 Next Calibration Date: 19-Oct-24 Operator: P.F.Yeung						
<b>CONDITIONS</b>							
Temperature (°C)	1005 30.0	Corrected Pressure (mm Hg)	753.8 303				
<b>CALIBRATION ORIFICE</b>							
Make: Model: Serial#:	TISCH TE-5025A 2454	Qstd Slope Qstd Intercept	2.07544 -0.03205				
<b>CALIBRATION</b>							
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION
18	5.8	5.8	11.6	1.637	53	52.36	Slope= 30.005
13	4.5	4.5	9.0	1.443	48	47.42	Intercept= 3.831
10	3.2	3.2	6.4	1.220	42	41.49	Corr. Coeff.= 0.9975
7	2.2	2.2	4.4	1.014	34	33.59	
5	1.4	1.3	2.7	0.798	28	27.66	

**Calulations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

$Q_{std}$  = standard flow rate  
 $IC$  = corrected chart response  
 $I$  = actual chart response  
 $m$  = calibrator  $Q_{std}$  slope  
 $b$  = calibrator  $Q_{std}$  intercept  
 $T_a$  = actual temperature during calibration (deg K)  
 $P_a$  = actual pressure during calibration (mm Hg)

**For subsequent calculation of sampler flow:**

$1/m((I)[\sqrt{298/T_{avg}}(P_{avg}/760)] - b)$

$m$  = sampler slope  
 $b$  = sampler intercept  
 $I$  = chart response  
 $T_{avg}$  = daily average temperature  
 $P_{avg}$  = daily average pressure

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM2 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Aug-24 Next Calibration Date: 19-Oct-24 Operator: P.F.Yeung							
<b>CONDITIONS</b>								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>30.0</td></tr> </table> <table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>303</td></tr> </table>	1005	30.0	753.8	303			
1005								
30.0								
753.8								
303								
<b>CALIBRATION ORIFICE</b>								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
<b>CALIBRATION</b>								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	5.5	5.5	11.0	1.594	52	51.37	Slope= 31.281	
13	4.1	4.1	8.2	1.379	47	46.43	Intercept= 2.566	
10	3.1	3.0	6.1	1.191	42	41.49	Corr. Coeff.= 0.9923	
7	2.3	2.2	4.5	1.025	34	33.59		
5	1.4	1.4	2.8	0.812	28	27.66		

**Calculations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

$Q_{std}$  = standard flow rate  
 $IC$  = corrected chart response  
 $I$  = actual chart response  
 $m$  = calibrator  $Q_{std}$  slope  
 $b$  = calibrator  $Q_{std}$  intercept  
 $T_a$  = actual temperature during calibration (deg K)  
 $P_a$  = actual pressure during calibration (mm Hg)

**For subsequent calculation of sampler flow:**

$1/m((I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$

$m$  = sampler slope  
 $b$  = sampler intercept  
 $I$  = chart response  
 $T_{avg}$  = daily average temperature  
 $P_{avg}$  = daily average pressure

Qstd (m³/min)	IC
0.812	27.66
1.025	33.59
1.191	41.49
1.379	46.43
1.594	51.37

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM3 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Aug-24 Next Calibration Date: 19-Oct-24 Operator: P.F.Yeung							
<b>CONDITIONS</b>								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>30.0</td></tr> </table>	1005	30.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>303</td></tr> </table>	753.8	303	
1005								
30.0								
753.8								
303								
<b>CALIBRATION ORIFICE</b>								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
<b>CALIBRATION</b>								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	5.2	5.2	10.4	1.551	57	56.31	Slope= 28.490	
13	4.1	4.1	8.2	1.379	52	51.37	Intercept= 11.913	
10	3.1	3.1	6.2	1.201	46	45.44	Corr. Coeff.= 0.9991	
7	2.0	2.0	4.0	0.967	40	39.52		
5	1.2	1.2	2.4	0.753	34	33.59		

**Calulations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

$Q_{std}$  = standard flow rate  
 $IC$  = corrected chart response  
 $I$  = actual chart response  
 $m$  = calibrator  $Q_{std}$  slope  
 $b$  = calibrator  $Q_{std}$  intercept  
 $T_a$  = actual temperature during calibration (deg K)  
 $P_a$  = actual pressure during calibration (mm Hg)

**For subsequent calculation of sampler flow:**

$1/m((I)[\sqrt{298/T_{avg}}(P_{avg}/760)] - b)$

$m$  = sampler slope  
 $b$  = sampler intercept  
 $I$  = chart response  
 $T_{avg}$  = daily average temperature  
 $P_{avg}$  = daily average pressure

**Flow Rate**

Qstd (m³/min)	IC
0.75	33.59
0.96	39.52
1.00	45.44
1.20	51.37
1.55	56.31

# TSP SAMPLER CALIBRATION CACULATION SPREADSHEET

Location ID: AM4 Name and Model : TISCH HVS Model TE-5170	Date of Calibration: 19-Aug-24 Next Calibration Date: 19-Oct-24 Operator: P.F.Yeung							
<b>CONDITIONS</b>								
Sea Level Pressure (hpa) Temperature (°C)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>1005</td></tr> <tr><td>30.0</td></tr> </table>	1005	30.0	Corrected Pressure (mm Hg) Temperature (K)	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>753.8</td></tr> <tr><td>303</td></tr> </table>	753.8	303	
1005								
30.0								
753.8								
303								
<b>CALIBRATION ORIFICE</b>								
Make: Model: Serial#:	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>TISCH</td></tr> <tr><td>TE-5025A</td></tr> <tr><td>2454</td></tr> </table>	TISCH	TE-5025A	2454	Qstd Slope Qstd Intercept	<table border="1" style="display: inline-table; vertical-align: middle;"> <tr><td>2.07544</td></tr> <tr><td>-0.03205</td></tr> </table>	2.07544	-0.03205
TISCH								
TE-5025A								
2454								
2.07544								
-0.03205								
<b>CALIBRATION</b>								
Plate No.	H2O(L) (in)	H2O(R) (in)	H2O (in)	Qstd (m3/min)	I (chart)	IC (corrected)	LINEAR REGRESSION	
18	6.3	6.4	12.7	1.712	56	55.32	Slope= 31.910	
13	4.8	4.9	9.7	1.498	51	50.38	Intercept= 1.594	
10	3.6	3.7	7.3	1.302	44	43.47	Corr. Coeff.= 0.9980	
7	2.4	2.5	4.9	1.069	36	35.57		
5	1.4	1.5	2.9	0.826	28	27.66		

**Calculations:**

$Q_{std} = 1/m[\sqrt{H_2O(P_a/P_{std})(T_{std}/T_a)} - b]$

$IC = I[\sqrt{P_a/P_{std}}(T_{std}/T_a)]$

$Q_{std}$  = standard flow rate  
 $IC$  = corrected chart response  
 $I$  = actual chart response  
 $m$  = calibrator  $Q_{std}$  slope  
 $b$  = calibrator  $Q_{std}$  intercept  
 $T_a$  = actual temperature during calibration (deg K)  
 $P_a$  = actual pressure during calibration (mm Hg)

**For subsequent calculation of sampler flow:**

$1/m((I)[\sqrt{(298/T_{avg})(P_{avg}/760)}] - b)$

$m$  = sampler slope  
 $b$  = sampler intercept  
 $I$  = chart response  
 $T_{avg}$  = daily average temperature  
 $P_{avg}$  = daily average pressure

