

ANNEX G2

CALIBRATION CERTIFICATES FOR LANDFILL GAS MONITORING EQUIPMENT



ALS Technichem (HK) Pty Ltd 11/F., Chung Shun Knitting Centre, 1 - 3 Wing Yip Street, Kwai Chung, N.T., Hong Kong T: +852 2610 1044 F: +852 2610 2021 www.alsglobal.com

CERTIFICATE OF ANALYSIS

CONTACT:MR IVAN LEUNGCLIENT:ALS TECHNICHEM (HK) PTY LTDADDRESS:11/F., CHUNG SHUN KNITTING CENTRE,
1-3 WING YIP STREET, KWAI CHUNG, N.T.

WORK ORDER: HK2437473

SUB BATCH:	0
LABORATORY:	HONG KONG
DATE RECEIVED:	14-Sep-2024
DATE OF ISSUE:	23-Sep-2024

SPECIFIC COMMENTS

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client. The performance of the equipment stated in this report is checked with independent reference material and results are compared against a calibrated secondary source.

The "Instrument Specification" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principles as practised by the laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

Equipment Type:	Landfill Gas Analyser
Service Nature:	Performance Check
Scope:	Carbon dioxide, Methane and Oxygen
Brand Name/ Model No.:	GA5000
Serial No./Equipment No.:	G507306 (HK1935)
Date of Calibration:	20 September, 2024

GENERAL COMMENTS

This report superseded any previous report(s) with same work order number.

Ms Anh Ngoc Huynh Senior Chemist - Organics

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

Page 1 of 2

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

Work Order: Sub-Batch: Client: Date of Issue:	HK2437473 0 ALS TECHNICHEM (HK) PTY LTD 23-Sep-2024
Equipment Type: Brand Name/ Model No.:	Landfill Gas Analyser GA5000
Serial No./ Equipment No.:	G507306 (HK1935)
Date of Calibration:	20 September, 2024



Next Calibration Date: 20 October, 2024

Parameters:

Methane

Calibrated Gas Standard, %	Monitor Readout, %	% error	Instrument Specification, %
0.0 (Nitrogen)	0.0	0.0	± 0.5
1.0	1.0	0.0	± 0.5
10.0	9.9	-0.1	± 0.5

Carbon Dioxide

Calibrated Gas Standard, %	Monitor Readout, %	% error	Instrument Specification, %
0.0 (Nitrogen)	0.0	0.0	± 0.5
1.0	1.0	0.0	± 0.5
10.1	9.9	-0.2	± 0.5

Oxygen

Calibrated Gas Standard, %	Monitor Readout, %	% error	Instrument Specification, %
0.0 (Nitrogen)	0.0	0.0	± 1.0
23.3	23.5	0.2	± 1.0

Ms Anh Ngoc Huynh Senior Chemist - Organics